

Corrigendum-etender for Supply, Installation of Membrane Test Skid

Tender Reference No: CFTRI/A3/79465A/20-21

Tender ID: 2020_CSIR_48958_1

Tender Title: Supply, Installation of **Membrane Test Skid**

Corrigendum Title: Corrigendum to specification (additional information-specification/drawing)/
Extension of date for submission of bid

Corrigendum Type: Before tender opening

Additional information/clarification

Clarification sought by one of the firm	Additional information/clarification
1) Drawing/details of existing Flatsheet Membrane holder	Drawing enclosed
2) existing holder is for Single layer membrane or cassette type membrane	Single layer membrane sheet to be used with existing holder

All other specification and Terms & Conditions of tender remains unaltered except for the following:

The last date for submission of bid stands extended up to 18th June 2020 upto 2.00 P.M

Bid opening (Technical) on line in etender.gov.in : 19th June 2020 at 2.30 P.M

Bidders are requested to note above corrigendum to extension of date and specification for bid submission and accordingly submit your offer with relevant documents .

Approved by


Signature

Dt. 10.6.2020

✓

Available Test Cell Details

Test cell dimensions 160*120 mm


Flat sheet membrane holder

Membrane size to be used 95 mm x 115 mm,

Effective membrane area 80 mm x 100 mm,

Max pressure 60 bar

Single layer membrane sheet to be used with existing holder



Mwanku
10/06/2020